



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Integrity Weighing Solutions LLC dba Lambert-Brown Scales
1101 E Louisville Street, Broken Arrow, OK 74012

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Mass, Force, and Weighing Devices and Mechanical Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

June 27, 2018

Issue Date:

December 03, 2024

Expiration Date:

February 28, 2027

Accreditation No.:

95375

Certificate No.:

L24-917

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Integrity Weighing Solutions LLC dba Lambert-Brown Scales

1101 E Louisville Street, Broken Arrow, OK 74012
 Contact Name: Jo Rothhammer Phone: 918-258-5515

Accreditation is granted to the facility to perform the following calibration:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Top Loaders ^{FO}	0.5 g to 20 kg (Resolution = 0.01 g)	$(1.16 \times 10^{-2} + 1.15 \times 10^{-3}Wt)$ kg	Class F Weights	NIST Handbook 44 WI- 101
Bench Scales ^{FO}	0.001 lb to 10 lb (Resolution = 0.001 lb)	$(1.2 \times 10^{-3} + 4.72 \times 10^{-5}Wt)$ lb		
	0.01 lb to 100 lb (Resolution = 0.01 lb)	$(1.16 \times 10^{-4} + 4.72 \times 10^{-5}Wt)$ lb		
Floor and Bench Scales ^{FO}	2 lb to 2 000 lb (Resolution = 0.5 lb)	$(5.77 \times 10^{-1} + 2.23 \times 10^{-5}Wt)$ lb		
Floor Scales ^{FO}	2 lb to 5 000 lb (Resolution = 0.5 lb)	$(5.77 \times 10^{-1} + 4.78 \times 10^{-5}Wt)$ lb		
	8 lb to 10 000 lb (Resolution = 2 lb)	$(2.31 + 2.73 \times 10^{-5}Wt)$ lb		
Pancake and Tank Scale ^{FO}	50 lb to 60 000 lb (Resolution = 10 lb)	$(11.55 + 3.2 \times 10^{-5}Wt)$ lb		
Truck/Wheel Scales ^{FO}	1 000 lb to 180 000 lb (Resolution = 20 lb)	$(23.05 + 4.46 \times 10^{-5}Wt)$ lb		
Class 1 Scales ^{FO}	20 mg to 320 g	$1.00 \times 10^{-4} + 2.37 \times 10^{-7}Wt)$ g	Class 1 Weights	

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Pressure Gauges ^{FO}	-14 psi to 100 psi	0.08 psi	Heise PTE 1 w/ HQS-2	ASME B40.1 WI 102
	100.5 psi to 1 000 psi	0.75 psi		
Pressure Devices ^{FO}	1 001 psi to 10 000 psi	2.9 psi		

- The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.



Certificate of Accreditation: Supplement

Integrity Weighing Solutions LLC dba Lambert-Brown Scales

1101 E Louisville Street, Broken Arrow, OK 74012
Contact Name: Jo Rothhammer Phone: 918-258-5515

Accreditation is granted to the facility to perform the following calibration:

2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations.
5. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.
6. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location